

DRRR GmbH · Bodmanstraße 4 · D-87435 Kempten

Mrs Danica Červinková  
VÚSAPL, a.s.  
poštový priečinok 59  
Novozámocká 178  
949 05 Nitra 5  
Slovakia

**DRRR-Proficiency Testing**  
**RVEP 200262 Total carbon emission**  
**VDA 277 - PV 3341**

Dear participant

Thank you very much for participating in our proficiency testing.

**Your lab code number is: 19**

We hope, that we will meet your laboratory in the next proficiency testing round again. Registrations for our program are possible any time using our online system (ODIN) or the registration forms on our homepage: [www.DRRR.de](http://www.DRRR.de)

New: The bilateral proficiency testing (bPT)!  
You are not satisfied with your laboratory performance? We will offer you our new bilateral proficiency testing. You can book and perform individually and flexibly the bPT within a ten week time period after this proficiency testing round. You will get your proof of performance as a z'-score in the form of a certificate. In case of arousing your interest, please get in contact with us.

For questions do not hesitate to contact us.

Yours sincerely

Thorsten Helbig

date: 14.08.2020

your contact persons:  
Thorsten Helbig  
Regina Wörz

fon: +49 (0)831/960 878-0  
fax: +49 (0)831/960 878-99  
e-mail: [info@DRRR.de](mailto:info@DRRR.de)

Deutsches Referenzbüro für Ringversuche und  
Referenzmaterialien GmbH  
Bodmanstraße 4 D-87435 Kempten  
fon: +49 (0)8 31/960 878-0  
fax: +49 (0)8 31/960 878-99  
e-Mail: [info@DRRR.de](mailto:info@DRRR.de)  
website: [www.DRRR.de](http://www.DRRR.de)

accounting details:  
VAT-no DE254613132 •TAX-no 127/124/32207  
Raiffeisenbank im Allgäuer Land  
bank code 733 692 64 •account no 102350  
IBAN DE 94733692640000102350  
BIC-Code: GENO DEF1DTA

place of business: Kempten  
commercial registry: HRB 9496  
local court: Kempten  
managing director:  
Dr. rer. nat. Ulrich Leist  
Thorsten Helbig M.Eng.

# CERTIFICATE FOR PARTICIPATION

## DRRR Proficiency Testing

The participant

VÚSAPL, a.s.  
poštový priečinok 59  
Novozámocká 178  
949 05 Nitra 5  
Slovakia

*Customer number:* 12626

*lab code number:* 19

gets for the above mentioned quantitative proficiency testing the following valuation:

**great success**

To the above mentioned proficiency testing the corresponding report contains the statistical evaluations.

Kempton, 14.08.2020



Thorsten Helbig

Valuation:

If the mean of z'-scores of all evaluated results is smaller than 2 or equal and single z'-scores are smaller than 2 or equal you will get the valuation: „great success“

If the mean of z'-scores of all evaluated results is smaller than 2 or equal and single z'-scores are higher than 2 you will get the valuation: „success“

If the mean of z'-scores of all evaluated results is higher than 2 you will get the valuation: „participated“

## Summary of your quantitative lab performance according to the reference method(s)

For your company

VÚSAPL, a.s.  
poštový priečinok 59

### SUMMARY

customer number: 12626  
lab code number: 19  
number of z'-scores: 2  
number of z'-scores > 2: 0

Valuation:

If the mean of z'-scores of a parameter is smaller than 2 or equal you will get the valuation:  
"You performed very well. Keep your performance."

If the mean of z'-scores of a parameter is smaller than 2 or equal but at least one single z'-score is higher than 2 you will get the valuation:  
"You performed well. In particular cases you showed potential for improvement."

If the mean z'-score of a parameter is higher than 2 you will get the valuation:  
"You showed potential for improvement of your performance."

parameter	number of z'-scores:	number of z'-scores > 2:	mean of z'-scores:	Valuation:
Total carbon emission	2	0	0,12	You performed very well. Keep your performance.